


Thermo Scientific Synchronis C8 HPLC Columns

Consistent, predictable separations,
Column after column, time after time

- Outstanding reproducibility
- Highly pure, high surface area silica
- Less hydrophobic than C18
- Double endcapped for extra surface coverage
- Highly inert towards basic compounds
- Rigorously tested to ensure quality

Specifications


Particle size	1.7 μm , 5 μm	Carbon load	10 %
Pore size	100 \AA	Endcapped	Y
Surface area	320 m^2/g	USP classification	L7
pH range	2 - 8		



Less retentive than C18

Synchronis™ C8 columns are less hydrophobic than the C18 and are therefore particularly useful where the lesser degree of hydrophobicity is needed in order to successfully retain compounds of interest. Synchronis C8 columns can also be used where it is desirable to elute compounds more quickly than with a C18 column.

Application: Fenopropfen (USP)



Column: Synchronis C8, 5 μm , 150mm x 4.6mm

Mobile phase: Acetonitrile:Water:
Phosphoric Acid (50:49.6:0.4)

Flow rate: 2.0 mL/min

Temperature: 30°C

Detection: 272 nm


Injection volume: 20 μL

1. Fenopropfen
2. Gemfibrozil

Parameter	USP Specification	Measured Fenopropfen (5 replicate injections)	Measured Gemfibrozil (5 replicate injections)
Resolution	> 8	–	17.6
Relative retention time	~ 0.5	0.48	–
Efficiency (N)	> 3000	9812	10254
Tailing factor	< 2	1.21	1.22
%RSD Retention time	< 2%	0.13%	0.14%
%RSD Peak area	< 2%	1.6%	1.8%

Description	Particle size	Length (mm)	2.1 mm ID	3 mm ID	4 mm ID	4.6 mm ID
Synchronis C8	1.7 µm	30	97202-032130	–	–	–
		50	97202-052130	97202-053030	–	97202-054630
		100	97202-102130	97202-103030	–	–
	5 µm	30	97205-032130	97205-033030	97205-034030	97205-034630
		50	97205-052130	97205-053030	97205-054030	97205-054630
		100	97205-102130	97205-103030	97205-104030	97205-104630
Drop-in guard cartridges (4/pk)	5 µm	10	97205-012101	97205-013001	97205-014001	–

Application: Uron Herbicides



Column: Synchronis C8, 5µm, 150mm x 4.6mm

Mobile phase: A: Water
B: Acetonitrile

Gradient: 35 to 60% B in 10 minutes

Flow rate: 1.0 mL/min

Temperature: 30°C

Detection: 240 nm

Injection volume.: 20 µL

1. Tebuthiuron
2. Metoxuron
3. Monuron
4. Chlorotoluron
5. Diuron
6. Linuron

Herbicide	RT (%RSD) (6 replicate injections)	Peak Area (%RSD) (6 replicate injections)	Peak Asymmetry
1 - Tebuthiuron	0.31	0.95	1.17
2 - Metoxuron	0.25	0.64	1.18
3 - Monuron	0.18	0.20	1.16
4 - Chlorotoluron	0.12	0.55	1.15
5 - Diuron	0.10	0.37	1.19
6 - Linuron	0.05	0.65	1.13

Consistent, predictable separations, Column after column, time after time

Synchronis HPLC columns are manufactured, packed and tested in ISO9000 accredited facilities. Each lot of silica is tested for the physical properties of the silica support and only released for production if it meets the stringent test specifications.

Each bonded lot of chromatographic packing material is rigorously tested for primary and secondary interactions with the bonded phase.

New, enhanced, automated packing methods drive consistency even further and every column is individually tested to ensure that it meets the required quality.

These extensive testing and quality control procedures ensure the delivery of a consistent product, column after column.

For more information, visit www.thermoscientific.com/Synchronis

©2010 Thermo Fisher Scientific Inc. All rights reserved. All trademarks are the property of Thermo Fisher Scientific Inc. and its subsidiaries. Specifications, terms and pricing are subject to change. Not all products are available in all countries. Please consult your local sales representative for details.

SSCCSYNCRONISC8 0810

USA and Canada +1 800 332 3331

France +33 (0) 3 88 67 53 20

Germany +49 6103 408 0

Switzerland +41 56 618 41 11

United Kingdom +44 (0) 1928 534 110

Japan +81 45 453 9220

China +86-21-68654588 or +86-10-84193588

India 1800 22 8374 (toll-free)

+91 22 6716 2200

All Other Enquiries +44 (0) 1928 534 050

Technical Support

North America +1 800 332 3331

Outside North America +44 (0) 1928 534 440

Thermo
SCIENTIFIC